

**Norway-Trondheim: Dental hand instrument**  
**OJ S 61/2016 26/03/2016**  
**Contract notice**  
**Supplies**

**Directive 2004/18/EC**

**Section I: Contracting authority**

---

**I.1. Name and addresses**

Official name: Sør-Trøndelag Fylkeskommune (Sør-Trøndelag County)

Postal address: Postboks 2350 Sluppen (Fylkeshuset E Skakkes g 14)

Town: Trondheim

Postal code: 7004

Country: Norway

For the attention of: Bjørn Pynten

E-mail: [bjorn.pynten@stfk.no](mailto:bjorn.pynten@stfk.no)

Telephone: +47 72812104

**Internet address(es):**

Electronic access to information: <http://permalink.mercell.com/56832126.aspx>

Electronic submission of tenders and requests to participate: <http://permalink.mercell.com/56832126.aspx>

**Additional information can be obtained from:**

Official name: Mercell Norge AS

Postal address: Karihaugveien 89

Town: Oslo

Postal code: 1086

Country: Norway

E-mail: [support@mercell.com](mailto:support@mercell.com)

Internet address: <http://permalink.mercell.com/56832126.aspx>

**Specifications and additional documents (including documents for competitive dialogue and a dynamic purchasing system) can be obtained from:**

Official name: Mercell Norge AS

Postal address: Karihaugveien 89

Town: Oslo

Postal code: 1086

Country: Norway

E-mail: [support@mercell.com](mailto:support@mercell.com)

Internet address: <http://permalink.mercell.com/56832126.aspx>

**Tenders or requests to participate must be submitted:** Official name: Mercell Norge AS

Postal address: Karihaugveien 89

Town: Oslo

Postal code: 1086

Country: Norway

E-mail: [support@mercell.com](mailto:support@mercell.com)

Internet address: <http://permalink.mercell.com/56832126.aspx>

**I.2. Type of the contracting authority**

Regional or local authority

### **I.3. Main activity**

General public services

### **I.4. Contract award on behalf of other contracting authorities**

The contracting authority is purchasing on behalf of other contracting authorities: no

## **Section II: Object of the contract**

---

### **II.1. Description**

#### **II.1.1. Title attributed to the contract by the contracting authority**

Dental equipment.

#### **II.1.2. Type of contract and place of performance or delivery**

Supplies

Purchase

Main site or place of performance: Sør-Trøndelag, Norway.

NUTS code NO061 Sør-Trøndelag

#### **II.1.3. Information about a framework agreement or a dynamic purchasing system**

The procurement involves the establishment of a framework agreement

#### **II.1.4. Information about framework agreement**

Framework agreement with a single operator

##### **Duration of the framework agreement**

Duration in years: 2

#### **II.1.5. Short description of the contract or purchase(s)**

Procurement of dental equipment.

The items requested are specified under the attachment for 'Partial Deliveries'. The assignment is divided into 5 groups, and it is possible to submit a tender offer on one or more groups. The tenderer must give an offer on all of the products in each group, and shall also be able to offer any extra equipment and accessories that naturally belong to each of the product groups, but that are not listed in the product lists. This means that the groups are evaluated individually and that the Contracting Authority may enter into contract with more than one tenderer.

The products that the Contracting Authority is requested offers on shall meet all requirements set by the Contracting Authority in the requirement specification. The requirement must be met if the tender offer can be considered and the documentation that the requirements are met shall be attached to the tender offer.

The Contracting Authority is Sør-Trøndelag fylkeskommune [county].

The Contracting Authority uses Mercell Sourcing Service. To express your interest and gain access to any documents, please click on the link below or copy and paste the link into your browser. [[<http://permalink.mercell.com/56832126.aspx>]]. Then follow the instructions on the website.

#### **II.1.6. CPV code(s)**

33131000 Dental hand instrument

#### **II.1.7. Information about the Government Procurement Agreement (GPA)**

The procurement is covered by the Government Procurement Agreement: no

#### **II.1.8. Lots**

This contract is divided into lots: yes  
Tenders may be submitted for one or more lots

**II.1.9. Information about variants**

Variants will be accepted: no

**II.2. Scope of the procurement**

**II.2.1. Total quantity or scope**

Estimated value excluding VAT: 8 000 000 NOK

**II.2.2. Information about options**

Options: no

**II.2.3. Information about renewals**

This contract is subject to renewal: no

**II.3. Duration of the contract or time limit for completion**

Start 9.5.2016. Completion 31.12.2017

Information about lots

Lot title: A. Unit, treatment chairs

**1) Short description**

Treatment chairs:

— Treatment chairs that can be adjusted to persons of all ages. set in different positions and have good support for both children and adults. The dental chair shall have soft upholstery and be comfortable to lie in for all types of patients, and should tolerate heavy patients (up to 150 kg). The surface shall have the necessary friction so that one can lie stable and dependably and shall be equipped with adjustable armrests. The chair should be equipped with armrests and a cover which tolerate soap washing and alcohol wipes. It should have opportunities for automatic tuning and a stop button. The armrests should be able to be moved out of the way in an easy manner.

**2) CPV code(s)**

33192400 Dental workstations

**3) Quantity or scope**

**4) Indication about different time frame or duration**

**5) Additional information about lots**

Lot title: B. Dental x-ray devices.

**1) Short description**

Dental x-ray devices:

— Wall mounted, ceiling mounted and unmounted x-ray devices must be offered.  
— The device shall be connected to the power grid with voltage 230/400 v.  
— The equipment must give low radiation doses and good radiation protection for patients and personnel. There should be an audible warning of exposure. The device shall be delivered with long rectangular cones and be usable both on children and adults.  
— All surface should be easily cleaned and disinfected. There should be minimum unhygienic openings in the surfaces. All surfaces should tolerate surface disinfection.  
— The x-ray equipment should be ergonomically good. The arms must be easy to move and at the same time be sturdy, precise and not move after the position is set.

— The control panel shall be fastened to the wall and with control of the exposure time. Electronic clock with control of variable kilo voltage and current strength. Set exposure time must match the actual exposure time.

**2) CPV code(s)**

33111500 Dental X-ray

**3) Quantity or scope**

**4) Indication about different time frame or duration**

**5) Additional information about lots**

Lot title: C. Contra angles, hand pieces and turbines.

**1) Short description**

Contra angles, hand pieces.

— A contra angle with exchange 1:1 for drill with diameter 2.35 mm should be able to be autoclaved up to 135 degrees Celsius and washed in a machine up to 95 degrees Celsius. It should have internal conveying of cooling to the drill and be equipped with a press button release of the drill.

— Red contra angle.

— Blue contra angel.

— Orange contra angle.

— Green contra angle.

— Lubricating machines for contra angles and turbines.

— The hand pieces must be able to be autoclaved in up to 135 degrees Celsius and washed in a machine with up to 95 degrees Celsius. Must have internal conveying of cooling to the drill and be equipped with fibre optic light.

— Micro motors that should have LED lighting.

— Technique motors.

Turbines:

— A turbine contra angle for use with standard FG drill (1.6 mm). Should be able to be autoclaved in up to 135 degrees Celsius and washed in a machine with up to 95 degrees Celsius. Shall have internal conveying of cooling to the drill and be equipped with push button release of the drill.

— Turbine connections should be able to delivered with and without LED light.

**2) CPV code(s)**

33130000 Dental and subspecialty instruments and devices, 42110000 Turbines and motors, 33136000 Rotary and abrasive instrument

**3) Quantity or scope**

**4) Indication about different time frame or duration**

**5) Additional information about lots**

Lot title: D. Compressors and suction motors.

**1) Short description**

Compressors:

— Should be able to connect to 220 V / 400 V alternating current.

— Should be equipped with dry air facility and deliver oil free and dry air.

— Should be equipped with minus 30 degrees pressure dew point.

— Should be equipped with pressure regulator five to ten bar.

— Price should be given on different capacity that corresponds with different clinic sizes of one to twelve units.

Suction motors:

— Should be able to connect to 220 V / 400 V alternating current.

— Price should be given on different capacity that corresponds with different clinic sizes from one to twelve offices/treatment units.

— Prices should be given on both complete suction facility and suction motor.

**2) CPV code(s)**

33130000 Dental and subspecialty instruments and devices

**3) Quantity or scope**

**4) Indication about different time frame or duration**

**5) Additional information about lots**

Lot title: E. Small equipment, autoclaves and dental washing machines.

**1) Short description**

Small equipment:

— No other requirement than that the tenderer must be able to offer the equipment, and it must meet medical technical requirements.

— Capsule mixer.

— Compound machine, impression materials.

— Gypsum vacuum mixer alginate blender.

— Vacuum press.

— Single gypsum trimmer with diamond disk.

— Double gypsum trimmer with diamond disk.

— Mobile dry suction for connecting to grinding box, with filter and dust bag.

— Grinding box..

— Welding machine/membrane welder.

— Mobile hardening lamp, LED.

— Ultrasonic bath.

Autoclaves:

— Autoclaves should satisfy the authority requirements for dental autoclaves. Connects to 230 V / 400 V. Class B autoclave.

— Must be able to be delivered with and without water rinser.

Dental washing machines:

— The dental washing machines shall satisfy all of the authority requirements for dental washing machines, including temperature and water quantity.

93 degrees for five minutes. Delivered with flat basket up and down. Can connect to 230 V / 400 V.

**2) CPV code(s)**

33130000 Dental and subspecialty instruments and devices, 33191110 Autoclaves

**3) Quantity or scope**

**4) Indication about different time frame or duration**

**5) Additional information about lots**

**Section III: Legal, economic, financial and technical information**

---

### **III.1. Conditions related to the contract**

#### **III.1.1. Deposits and guarantees required**

#### **III.1.2. Main financing conditions and payment arrangements and/or reference to the relevant provisions governing them**

#### **III.1.3. Legal form to be taken by the group of economic operators to whom the contract is to be awarded**

#### **III.1.4. Contract performance conditions**

The performance of the contract is subject to particular conditions: no

### **III.2. Conditions for participation**

#### **III.2.1. Suitability to pursue the professional activity, including requirements relating to enrolment on professional or trade registers**

List and brief description of conditions: Tax certificate Tax and VAT — Requirement: The tenderer shall have their affairs in order in relation to payment of Taxes. Documented by: A certificate for Tax and VAT. The certificate must not be more than six months old when calculated from the deadline for submission of application. Any arrears or other irregularities must be justified. If the tenderer is planning to engage sub-suppliers, certificates for them must also be submitted. In Norway the certificates are issued by The Norwegian Tax Administration. Tenderers from other countries shall enclose equivalent certificates from the country where the tenderer is established.

Company Registration Certificate — Requirement: The tenderer must be a legally established company. Documentation: Norwegian companies: Company Registration Certificate. Foreign companies: Certificate verifying that the company is registered in a trade index or a Register of Business Enterprises, as required by the country where the tenderer is established.

#### **III.2.2. Economic and financial ability**

List and brief description of conditions: Annual accounts — Requirement: The tenderer must have the financial capacity to implement the assignment. Documentation: Annual accounts, including the board's annual report and auditor's report.

Credit appraisal — Requirement: The tenderer must have the sufficient financial capacity to implement the assignment. Documentation: Credit appraisal/rating, not older than six months, which is based on the last known accounting figures. The rating must be carried out by a publicly approved credit rating institution.

#### **III.2.3. Technical and professional ability**

List and brief description of conditions:

Experience — Requirement: The tenderer shall have experience from a minimum of three similar assignments during the last 3 years. Here, 'similar assignments' means assignments for large government organisations. Documentation: Description of the tenderer's three most relevant assignments during the last 3 years. The description must include a statement of the contract's value, date and recipient (name, telephone number and e-mail address). The tenderers must substantiate the assignments' relevance through the description.

Manpower and equipment — Requirement: The tenderer shall have sufficient implementation ability and capacity. Documentation: A statement of the tenderer's average workforce during the last 3 years and the equipment, machinery, tools, material or technical equipment that are available to the tenderer in connection with implementation of the agreement/assignment.

Quality assurance system/management system — Requirement: The services that are to be

provided must have a good and well-functioning quality assurance system. Documentation: A statement regarding the tenderer's quality assurance system/quality management system or a copy of the system certificate issued by accredited certification bodies, or similar documentation.

Social Contract Terms: Requirement: Quality assurance system shall ensure that the tenderer is suited to meet the social contract terms (see 'General contract terms point 7.5.1' and the attached form 'Qualification Requirements — social contract terms') Documented by: Attached form (Qualification requirement — social contract terms) shall be answered with the requested documentation.

A report on the tenderer's environmental competence and procedures that document that the requirement is met regarding environmental conditions. If this is described in the organisation's quality or environmental management system in accordance with ISO 9001, and ISO 14001, or other third party verified systems, it is sufficient to submit a copy of the current certificate.

#### **III.2.4. Information about reserved contracts**

#### **III.3. Conditions specific to services contracts**

##### **III.3.1. Information about a particular profession**

##### **III.3.2. Information about staff responsible for the performance of the contract**

### **Section IV: Procedure**

---

#### **IV.1. Type of procedure**

##### **IV.1.1. Type of procedure**

Open

##### **IV.1.2. Information about the limits on the number of candidates to be invited**

##### **IV.1.3. Information about reduction of the number of solutions or tenders during negotiation or dialogue**

#### **IV.2. Award criteria**

##### **IV.2.1. Award criteria**

The most economically advantageous tender in terms of Price is not the only award criterion and all criteria are stated only in the procurement documents

##### **IV.2.2. Information about electronic auction**

An electronic auction will be used: no

#### **IV.3. Administrative information**

##### **IV.3.1. File reference number attributed by the contracting authority**

201605515

##### **IV.3.2. Previous publication concerning this procedure**

no

##### **IV.3.3. Conditions for obtaining specifications and additional documents or descriptive document**

Payable documents: no

**IV.3.4. Time limit for receipt of tenders or requests to participate**

27.4.2016 - 09:00

**IV.3.5. Estimated date of dispatch of invitations to tender or to participate to selected candidates**

**IV.3.6. Languages in which tenders or requests to participate may be submitted**

Other: Norwegian.

**IV.3.7. Minimum time frame during which the tenderer must maintain the tender**

until: 23.6.2016

**IV.3.8. Conditions for opening of tenders**

Date: 27.4.2016 - 12:00

Place:

Trondheim

Persons authorised to be present at the opening of tenders: no

**Section VI: Complementary information**

---

**VI.1. Information about recurrence**

This is a recurrent procurement: no

**VI.2. Information about European Union funds**

The procurement is related to a project and/or programme financed by European Union funds:  
no

**VI.3. Additional information**

**VI.4. Procedures for review**

**VI.4.1. Review body**

**VI.4.2. Review procedure**

**VI.4.3. Service from which information about the review procedure may be obtained**

**VI.5. Date of dispatch of this notice**

23.3.2016