

Germany-Gelsenkirchen: X-ray devices

OJ S 177/2021 13/09/2021

Contract notice

Supplies

Legal Basis:

Directive 2014/24/EU

Section I: Contracting authority

I.1. Name and addresses

Official name: Westfälische Hochschule

Postal address: Neidenburger Straße 43

Town: Gelsenkirchen

NUTS code: DEA32 Gelsenkirchen, Kreisfreie Stadt

Postal code: 45897

Country: Germany

Contact person: Dezernat I.3 Einkauf

E-mail: ausschreibung@w-hs.de

Internet address(es):

Main address: www.w-hs.de

I.3. Communication

The procurement documents are available for unrestricted and full direct access, free of charge, at: <https://www.evergabe.nrw.de/VMPSatellite/notice/CXPNY6LDE7G/documents>

Additional information can be obtained from the abovementioned address

Tenders or requests to participate must be submitted electronically via: <https://www.evergabe.nrw.de/VMPSatellite/notice/CXPNY6LDE7G>

I.4. Type of the contracting authority

Body governed by public law

I.5. Main activity

Education

Section II: Object

II.1. Scope of the procurement

II.1.1. Title

Mikro-Röntgen-CT

Reference number: 10/2021 EU

II.1.2. Main CPV code

33111000 X-ray devices

II.1.3. Type of contract

Supplies

II.1.4. Short description

Gesucht wird ein μ CT der neuesten Bauart als Tischgerät zum Einsatz in Forschung und Lehre an der Westfälischen Hochschule.

II.1.5. Estimated total value

Value excluding VAT: 600 000,00 EUR

II.1.6. Information about lots

This contract is divided into lots: no

II.2. Description

II.2.2. Additional CPV code(s)

33111000 X-ray devices

II.2.3. Place of performance

NUTS code: DEA34 Borken

Main site or place of performance: Westfälische Hochschule Münsterstr. 265 46397 Bocholt

II.2.4. Description of the procurement

Gesucht wird ein μ CT der neuesten Bauart als Tischgerät zum Einsatz in Forschung und Lehre an der Westfälischen Hochschule.

Das Gerät soll weitgehend wartungsfrei sein und Um- und Nachrüstungen von Primärkomponenten müssen möglich sein. Da das Gerät sowohl in der Forschung, als auch in der Lehre eingesetzt wird, müssen die Strahlenschutz und Sicherheitseinrichtungen den Anforderungen an ein Vollschutzsystem gemäß deutscher RöV ($< 1 \mu\text{Sv/h}$) entsprechen. Der Probenraum soll bei geöffneter Tür von vorne, hinten und oben zugänglich sein. Die Probenkammer soll mindestens 400 mm Höhe haben (Helix CT). Aufstellung erfolgt in einem üblichen Labor ohne speziellen Industrieboden, das Gewicht darf daher 600 kg nicht überschreiten, d.h. muss noch hubwagentauglich sein und für Türen bis 90 cm Breite geeignet sein. Die Wärmelast darf 500 W nicht übersteigen. Für die Installation muss die Stromversorgung 1-phasig, abgesichert mit 16 A sein. Es wird eine 3D-Röntgenanlage mit 130 kV und Vollschutzgehäuse als Tischgerät, inklusive Unterbaukonstruktion erwartet. Die Auflösung soll in einem Bereich von unter $5\mu\text{m}$ liegen. Eine verbesserte Grauwertauflösung soll erreicht werden durch einen zusätzlichen photonenzählenden Detektor.

II.2.5. Award criteria

Criteria below

Price

II.2.6. Estimated value

Value excluding VAT: 600 000,00 EUR

II.2.7. Duration of the contract, framework agreement or dynamic purchasing system

Start: 01/10/2021

This contract is subject to renewal: no

II.2.10. Information about variants

Variants will be accepted: no

II.2.11. Information about options

Options: no

II.2.13.

Information about European Union funds

The procurement is related to a project and/or programme financed by European Union funds:
no

II.2.14. Additional information

Section IV: Procedure

IV.1. Description

IV.1.1. Type of procedure

Open procedure

IV.1.3. Information about a framework agreement or a dynamic purchasing system

IV.1.8. Information about the Government Procurement Agreement (GPA)

The procurement is covered by the Government Procurement Agreement: no

IV.2. Administrative information

IV.2.2. Time limit for receipt of tenders or requests to participate

Date: 15/10/2021 Local time: 10:00

IV.2.3. Estimated date of dispatch of invitations to tender or to participate to selected candidates

IV.2.4. Languages in which tenders or requests to participate may be submitted

German

IV.2.6. Minimum time frame during which the tenderer must maintain the tender

Tender must be valid until: 31/12/2021

IV.2.7. Conditions for opening of tenders

Date: 15/10/2021 Local time: 10:00

Section VI: Complementary information

VI.1. Information about recurrence

This is a recurrent procurement: no

VI.2. Information about electronic workflows

Electronic ordering will be used

Electronic invoicing will be accepted

Electronic payment will be used

VI.3. Additional information

Bekanntmachungs-ID: CXPNY6LDE7G

VI.4. Procedures for review

VI.4.1. Review body

Official name: Vergabekammer Westfalen

Postal address: Albrecht-Thaer-Straße 9

Town: Münster

Postal code: 48147

Country: Germany

VI.5. Date of dispatch of this notice

08/09/2021